

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000337

Service date: 16.02.2026

Valid until: 16.02.2027

Result: **Approved**

Equipment details

Customer: Otech AS
Producer: Ashford Instruments
Serial no: 420706-4
Customer tag: Diver 1

Location: 10Ft
Model: Pneumo 0-76M
Build year: —
Service interval: 12 months

Calibration Instrument

Input basis: msw

Accuracy (%FS): 0.25

Range: 0 + 76 m

Bar → msw factor: 1

Full scale: 76 msw

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	5.012	0.012	0.0158	PASS
2	10.0	10.012	0.012	0.0158	PASS
3	20.0	20.033	0.033	0.0434	PASS
4	30.0	30.03	0.030	0.0395	PASS
5	40.0	40.031	0.031	0.0408	PASS
6	50.0	50.042	0.042	0.0553	PASS

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

