

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000345

Service date: 18.02.2026

Valid until: 18.02.2027

Result: **Approved**

Equipment details

Customer: Otech AS
Producer: Novasub
Serial no: 371129007
Customer tag: Diver 3

Location: Transporter
Model: Pneumo 0-60M
Build year: 2024
Service interval: 12 months

Calibration Instrument

Input basis: msw

Accuracy (%FS): 0.25

Range: 0 + 60 m

Bar → msw factor: 1

Full scale: 60 msw

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	5.0	0.000	0.0000	PASS
2	10.0	10.02	0.020	0.0333	PASS
3	20.0	20.05	0.050	0.0833	PASS
4	30.0	30.006	0.006	0.0100	PASS
5	40.0	40.01	0.010	0.0167	PASS
6	50.0	50.08	0.080	0.1333	PASS

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

