

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000454

Service date: 23.02.2026

Valid until: 23.02.2027

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: Gauge
Serial no: 006 Panel
Customer tag: HP divepanel

Location: Mjø
Model: Gauge HP
Build year: —
Service interval: 12 months

Calibration Instrument

Input basis: bar

Range: 400 bar

Full scale: 400 bar

Accuracy (%FS): 1.6

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS
Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (bar)	Reference (bar)	Dev (bar)	%FS	Result
1	50.0	51.0	1.000	0.2500	PASS
2	100.0	101.0	1.000	0.2500	PASS
3	150.0	152.4	2.400	0.6000	PASS
4	200.0	202.1	2.100	0.5250	PASS
5	250.0	253.1	3.100	0.7750	PASS
6	0.0	0.0	0.000	0.0000	PASS

Remarks

new 2025

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

