

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000562

Service date: 18.03.2026

Valid until: 18.03.2027

Result: **Approved**

Equipment details

Customer: Valen marina AS

Producer: 3D instruments

Serial no: D2green

Customer tag: Diver 2 Chambers panel

Location: DS1

Model: Pneumo Gauge

Build year: —

Service interval: 12 months

Calibration Instrument

Input basis: msw

Accuracy (%FS): 0.25

Range: 0 + 60 m

Bar → msw factor: 10

Full scale: 60 msw

Calibration Instrument: —

Calibration Instrument verified against: —

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	0.511	0.110	0.1833	PASS
2	10.0	1.01	0.100	0.1667	PASS
3	20.0	2.0129	0.129	0.2150	PASS
4	30.0	3.0148	0.148	0.2467	PASS
5	50.0	5.0038	0.038	0.0633	PASS

Remarks

No serial number on dial. Chambers panel

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

