

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000565

Service date: 18.03.2026

Valid until: 18.03.2027

Result: **Approved**

Equipment details

Customer: Valen marina AS
Producer: 3D instruments
Serial no: 99839-19-21
Customer tag: Diver 1 Blue

Location: Poseidon
Model: Pneumo Gauge
Build year: 2021
Service interval: 12 months

Calibration Instrument

Input basis: msw
Accuracy (%FS): 0.25

Range: 0 + 60 m
Bar → msw factor: 10

Full scale: 60 msw

Calibration Instrument: —
Calibration Instrument verified against: —

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	0.513	0.130	0.2167	PASS
2	10.0	1.011	0.110	0.1833	PASS
3	20.0	2.053	0.530	0.8833	FAIL
4	30.0	3.036	0.360	0.6000	FAIL
5	50.0	5.065	0.650	1.0833	FAIL

Remarks

Gauge are out of Range, Not in use to measure depth. Only for floating umbilical or liftbag

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

