

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000734

Service date: 28.05.2026

Valid until: 28.05.2027

Result: **Approved**

Equipment details

Customer: Valen marina AS
Producer: Ashford Instruments
Serial no: 102667 / D2
Customer tag: Diver 2

Location: Falken
Model: Digital Pneumo
Build year: —
Service interval: 12 months

Calibration Instrument

Input basis: msw

Range: 0 + 60 m

Full scale: 60 msw

Accuracy (%FS): 0.25

Bar → msw factor: 10

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +/-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	0.504	0.040	0.0667	PASS
2	10.0	1.007	0.070	0.1167	PASS
3	20.0	2.011	0.110	0.1833	PASS
4	30.0	3.017	0.170	0.2833	FAIL
5	50.0	5.027	0.270	0.4500	FAIL

Technician

Serviced by: Daniel Kristensen

Signature

