

HYPERBARIC CHAMBER SERVICE CERTIFICATE

DSC-2026-000261

Service date: 28.01.2026

Valid to: 20.10.2026

Result: **Approved**

Equipment details

Customer: Amundsen Diving AS
Producer: Haux
Serial no: 200604
Service interval: 12 months

Location: Scuba
Model: 1800MM
Build year: 2006

Pressure

Pressure reducing valve	20.0	Pass
Main chamber overpressure valve	4.7	Pass
Entry lock overpressure valve	—	Pass
Pipe maximum working pressure	—	Pass

Pressure test
Pressure: — bar
Time: — min
Descent: — min
Ascent: — min

Hydro
Last hydrostatic test:
20.10.2016
Next hydrostatic test:
20.10.2026
Hydrotested: No
Testrange: —

Checklist

General condition	Pass	Communication – main chamber	Pass
Entry lock inlet valve	Pass	Communication – Entry lock	Pass
Entry lock outlet valve	Pass	Gasket – medical lock	Pass
Main chamber inlet valve	Pass	Gasket – Entry lock door	Pass
Main chamber outlet valve	Pass	Gasket – main chamber door	Pass
Depth alarm – audible signal	-	Window/viewport New 2017	Pass
Depth gauge Calibrated	Pass	Valves & Piping	Pass
High pressure gauge (HP)	Pass	Lightening	Pass
Low pressure gauge (LP)	Pass	Cameras	-
BIBS / O ₂ mask	Pass	Surface treatment inside/outside	Pass
LP bank	-	Equalizer	Pass
HP Bank	Pass	Drainage	Pass
O ₂ needle valves internal / external	Pass		

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

