

ANALYZER CALIBRATION

Certificate no: DSC-2026-000490

Service date: 25.02.2026

Valid until: 25.02.2027

Result: **Approved**

Equipment details

Customer: Otech AS
Producer: Analytical industries
Serial no: 407257457
Service interval: 12 months

Location: CCB
Model: All-3000
Build year: -

NOTE

We confirm that Above analyzer has been Tested and calibrated with 100% oxygen and 20,9% air.
Oxygen sensor is more then 1 year and recommended to be replaced.

100% o2 - Adjusted to 100%
20,9% air - Adjusted to 20,9%

Changed battery and Ordered new sensor that will be supplied to customer

Technician

Serviced by: Daniel Kristensen

Signature

