

INTERSPIRO DIVATOR SET

Certificate no: DSC-2026-000198

Service date: 10.12.2025

Valid until: 10.12.2026

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS

Producer: Interspiro

Serial no: P1000204 ,9932045, 043FF6A16A2681

Service interval: 12 months

Location: Kammerkontainer

Model: Divator MKII

Build year: -

NOTE

Interspiro Divator MKII set- Mask,Regulator and 1.stage. Backup

This is a certificate confirmation, the documentation has a separate file.

Technician

Serviced by: Daniel Kristensen

Signature

Electronically signed by DS Bergen AS



test report

test report: 375

device no.: 043FF6A16A2681

type of test: Yearly test

test system: PRO

3102.24.325

workshop: DS BERGEN AS		person:	Daniel Kristensen
street: Storebotn 55G		phone:	91649162
place: 5309 Kleppestø		fax:	
		email:	Daniel@dsbergen.no
customer: Sjøentreprenøren AS		person:	Ingmar Aune
street:		phone:	
place: Bergen		fax:	
		email:	ingmar.aune@ncc.no
device data:	deadlines:	device-info:	service:
device no: 043FF6A16A2681	1 y test 10.12.2026 1 y srv 10.12.2026		- 1 year service - cleaning and disinfection
type: Interspiro Divator NP			
class: mask NP			
location: Kammerkontainer			

Reserve sett Divator Komplet.

tests/measurements	unit	min	max	actual	result
visual test (0 = ok) input mask: (0 = classic, > 0 = extended/list...)			0	0	OK
mask fitted approx. 6 sec head expansion Mask volume 0,2 l				0	OK
leakage test w/ open exhalation valve 0,25 min at p=-10 mbar	mbar		7,5	3,5	OK

test executed: 10.12.2025 21:28
by: Daniel Kristensen

result:

test signed: 10.12.2025 21:28
by: Daniel Kristensen

device ready for use

Lab.Net Profi V: 2024.0.0.0

next test on: 10.12.2026

test report

test report: 375
device no.: 9932045
type of test: Yearly test
test system: PRO

3102.24.325

workshop: DS BERGEN AS		person:	Daniel Kristensen
street: Storebotn 55G		phone:	91649162
place: 5309 Kleppestø		fax:	
		email:	Daniel@dsbergen.no
customer: Sjøentreprenøren AS		person:	Ingmar Aune
street:		phone:	
place: Bergen		fax:	
		email:	ingmar.aune@ncc.no
device data:	deadlines:	device-info:	service:
device no: 9932045	1 y test 10.12.2026 1 y srv 10.12.2026		1 year service cleaning and disinfection
type: Interspiro BV Divator NP			
class: demand valve NP			
location: Kammerkontainer			

Reserve sett Divator Komplet.

tests/measurements	unit	min	max	actual	result
visual test (0 = ok) input mask: (0 = classic, > 0 = extended/list...)			0	0	OK
demand valve opening pressure at 10 l/min suction	mbar	-4	-0,5	-1,8	OK
inhalation resistance at 40 x 2,5 l/min, 5 hubs HP .>. bar	mbar	-8	-0,1	-3,1	OK
exhalation valve opening pressure at 10 l/min suction	mbar	2	5	2,2	OK
safety valve opening pressure	bar	12	20	13,4	OK
pressure release of test equipment (HP, MP and LP)				0	OK

test executed: 10.12.2025 21:28
by: Daniel Kristensen
test signed: 10.12.2025 21:28
by: Daniel Kristensen

result:

device ready for use

Lab.Net Profi V: 2024.0.0.0

next test on: 10.12.2026

test report

test report: 375
device no.: P1000204
type of test: Yearly test
test system: PRO

3102.24.325

workshop: DS BERGEN AS		person:	Daniel Kristensen
street: Storebotn 55G		phone:	91649162
place: 5309 Kleppestø		fax:	
		email:	Daniel@dsbergen.no
customer: Sjøentreprenøren AS		person:	Ingmar Aune
street:		phone:	
place: Bergen		fax:	
		email:	ingmar.aune@ncc.no
device data:	deadlines:	device-info:	service:
device no: P1000204	1 y test 10.12.2026 1 y srv 10.12.2026		1 year service cleaning and disinfection
type: Poseidon Cyklon			
class: diving system basic			
location: Kammerkontainer			

Reserve sett Divator Komplet.

tests/measurements	unit	min	max	actual	result
visual test (0 = ok) input mask: (0 = classic, > 0 = extended/list...)			0	0	OK
HP leakage at max pressure after 0,5 min	bar	-5		-1	OK
demand valve opening pressure	bar	12	21	13,4	OK
static medium pressure at 150 bar cylinder pressure	bar	8	10,5	8,4	OK
MP leakage after 0,5 min at 190 bar lmt. 1 0=rel.1=abs. pulse	bar		10,5	8,7	OK
pressure comparison at 190 bar	bar	-10	10	10	OK
pressure comparison at 60 bar	bar	-10	10	7	OK
pressure release of test equipment (HP, MP and LP)				0	OK

test executed: 10.12.2025 21:28
by: Daniel Kristensen

result:

test signed: 10.12.2025 21:28
by: Daniel Kristensen

device ready for use

Lab.Net Profi V: 2024.0.0.0

next test on: 10.12.2026