

ANALYZER CALIBRATION

Certificate no: DSC-2026-000575

Service date: 19.03.2026

Valid until: 19.03.2027

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: Deox
Serial no: 0102690011
Service interval: 12 months

Location: Mjø
Model: Multigas
Build year: 2023

NOTE

DEOX multigas analyzer calibration.

Analyzer is tested with known gas.

O₂ - 21,2 %

CO₂ - 370 PPM

CO - 0.0 PPM

VOC - 0,2 MG/M³

H₂O - 0,0 MG/M³

O₂ - 98,1 %

CO₂ - 0 PPM

CO - 0.0 PPM

VOC - 0,0 MG/M³

H₂O - 0,0 MG/M³

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

