

OXYGEN REGULATOR

Certificate no: DSC-2026-000656

Service date: 15.04.2026

Valid until: 15.04.2027

Result: **Approved**

Equipment details

Customer: Kystarbeid AS
Producer: Zenon Regulator
Serial no: KA2014
Service interval: 12 months

Location: Chamber Container
Model: Oxygen regulator
Build year: -

NOTE

Oxygen regulator ZENON 0-10Bar Tested and approved.
OPV opens at 15.5 Bar

Oxygen cleaned and ready to use

Technician

Serviced by: Daniel Kristensen

Signature

