

ANALYZER CALIBRATION

Certificate no: DSC-2026-000667

Service date: 17.04.2026

Valid until: 17.04.2027

Result: **Approved**

Equipment details

Customer: Otech AS
Producer: Divesoft
Serial no: DS462
Service interval: 12 months

Location: Chamber
Model: Analyzer HE/O2
Build year: 2023

NOTE

Calibration of Divesoft HE and O2 analyzer,

Before

O2 - 21,3 %

O2 100,5 %

HE 0,5 %

HE 101,1 %

After

O2 - 20,9 %

O2 99,9 %

HE 0,1 %

HE 100,1 %

Oxygen sensor replaced 08/2025 D05R Molex sensor. Next change 08/2026

Technician

Serviced by: Daniel Kristensen

Signature

