


workshop: DS BERGEN AS		person:	Daniel Kristensen
street: Storebotn 55G		phone:	91649162
place: 5309 Klippeøstø		fax:	
		email:	Daniel@dsbergen.no
customer: IMC Di v i n g AS		person:	Jarle
street: Kristiansholm 8		phone:	
place: 5035 Bergen		fax:	
		email:	jarle.johansen@mc-di v i n g.no
device data:	deadlines:	device-info:	service:
device no: M2113017	1 y test 16.04.2027		- 1 year service - cleaning and disinfection
	1 y srv 16.04.2027		
M2113017	6 y srv 16.09.2031		
type: Interspiro Divator PP			
class: mask PP			
location: 137			

nr 137 maske med kirby blokk til aga maske

tests/measurements	unit	min	max	actual	result
visual test (0 = ok) input mask: (0 = classic, > 0 = extended/list...)			0	0	OK
mask fitted approx. 6 sec head expansion Mask volume 0,2 l				0	OK
leakage test w/ open exhalation valve 0,25 min at p=-10 nbar	nbar		7,5	0,9	OK

test executed: 16.04.2026 11:45
 by: Daniel Kristensen
 test signed: 16.04.2026 11:45
 by: Daniel Kristensen
 LabNet V: 2025.12.2.0
 next test on: **16.04.2027**

result:
device ready for use

workshop: DS BERGEN AS		person:	Daniel Kristensen
street: Storebotn 55G		phone:	91649162
place: 5309 Klippeøstø		fax:	
		email:	Daniel@dsbergen.no
customer: MC Diving AS		person:	Jarle
street: Kristiansholm 8		phone:	
place: 5035 Bergen		fax:	
		email:	jarle.johansen@mc-diving.no
device data:	deadlines:	device-info:	service:
device no: B2241213	1 y test 16.04.2027 1 y srv 16.04.2027 6 y srv 16.09.2031		1 year service cleaning and disinfection
type: Interspiro BV Divator PP			
class: demand valve overpressure			
location: 137			

nr 137 maske med kirby blokk til aga maske

tests/measurements	unit	min	max	actual	result
visual test (0 = ok) input mask: (0 = classic, > 0 = extended/list...)			0	0	OK
demand valve activation pressure at 10 l/min suction	nbar	-25	-0,1	-7,4	OK
static positive pressure at bar cylinder pressure, suction 0 0=yes 1=no	nbar	2	5	4,4	OK
inhalation resistance at 40 x 2,5 l/min 5 hubs HP .> bar	nbar	0,2	10	1	OK
pressure release of test equipment (HP, MP and LP)				0	OK

test executed: 16.04.2026 11:45
 by: Daniel Kristensen
 test signed: 16.04.2026 11:45
 by: Daniel Kristensen

result:

device ready for use

LabNet V: 2025.12.2.0

next test on: 16.04.2027