


<b>workshop:</b> DS BERGEN AS		person:	Daniel Kristensen
street: Storebotn 55G		phone:	91649162
place: 5309 Klippepestø		fax:	
		email:	Daniel@dsbergen.no
<b>customer:</b> Havforskningsinstituttet		person:	Børge Alfstad
street: Nordnesgaten 50		phone:	93201295
place: 5005 Bergen		fax:	
		email:	boerge.alfstad@hi.no
<b>device data:</b>	<b>deadlines:</b>	<b>device-info:</b>	<b>service:</b>
device no: <b>M2105174</b>  M2105174	1 y test 04.02.2027 1 y srv 04.02.2027		- 1 year service - cleaning and disinfection
type: Interspiro Divator PP			
class: mask PP			
location: Bergen			

tests/measurements	unit	min	max	actual	result
visual test (0 = ok input mask: (0 = classic, > 0 = extended/list...))			0	0	OK
mask fitted approx. 6 sec head expansion Mask volume 0,2 l				0	OK
leakage test w/ open exhalation valve 0,25 min at p=-10 nbar	nbar		7,5	1,3	OK

test executed: 04.02.2026 20:55  
 by: Daniel Kristensen  
 test signed: 04.02.2026 20:55  
 by: Daniel Kristensen  
 LabNet V: 2025.12.2.0  
 next test on: 04.02.2027

**result:**  
 device ready for use

<b>workshop:</b> DS BERGEN AS		person:	Daniel Kristensen
street: Storebotn 55G		phone:	91649162
place: 5309 Klippestø		fax:	
		email:	Daniel@dsbergen.no
<b>customer:</b> Havforskningsinstituttet		person:	Børge Alfstad
street: Nordnesgaten 50		phone:	93201295
place: 5005 Bergen		fax:	
		email:	boerge.alfstad@hi.no
<b>device data:</b>	<b>deadlines:</b>	<b>device-info:</b>	<b>service:</b>
device no: <b>1633081</b>	1 y test 04.02.2027 1 y srv 04.02.2027 6 y srv 13.03.2031		1 year service cleaning and disinfection
type: Interspiro BV Divator PP			
class: demand valve overpressure			
location: Bergen			

tests/measurements	unit	min	max	actual	result
visual test (0 = ok) input mask: (0 = classic, > 0 = extended/list...)			0	0	OK
demand valve activation pressure at 10 l/min suction	mbar	-25	-0,1	-17,8	OK
static positive pressure at bar cylinder pressure, suction 0 0=yes 1=no	mbar	2	5	4,1	OK
inhalation resistance at 40 x 2,5 l/min 5 hubs HP .> bar	mbar	0,2	10	0,2	OK
pressure release of test equipment (HP, MP and LP)				0	OK

test executed: 04.02.2026 20:55  
 by: Daniel Kristensen  
 test signed: 04.02.2026 20:55  
 by: Daniel Kristensen  
 LabNet V: 2025.12.2.0  
 next test on: **04.02.2027**

**result:**  
**device ready for use**