

OXYBOX SERVICE CERTIFICATE

DSC-2026-000290

Service date: 30.01.2026

Valid until: 30.01.2027

Result: **Approved**

Equipment details

Customer: Amundsen Diving AS
Producer: CBM
Serial no: 12785
Service interval: 12 months

Location: Scuba
Model: Oxybox
Build year: —

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	01.03.2025	01.03.2035	—	-
1st stage SN 031131793A	2nd stage SN 20070717	Cylinder SN 25AG0o2 003UT		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Comment: Changed	
		Flow test regulator	Pass

Notes

New cylinder, Oxygen cleaned and filled up to 200Bar

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen