

OXYBOX SERVICE CERTIFICATE

DSC-2026-000351

Service date: 18.02.2026

Valid until: 18.02.2027

Result: **Approved**

Equipment details

Customer: Otech AS
Producer: CBM
Serial no: 2072
Service interval: 12 months

Location: Transporter
Model: Oxybox
Build year: 2015

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	01.03.2025	01.03.2035	—	-
1st stage SN 15-06/339	2nd stage SN -	Cylinder SN 25AG062 013UT		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Flow test regulator	Pass

Notes

New cylinder 2026.

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

