

OXYBOX SERVICE CERTIFICATE

DSC-2026-000373

Service date: 21.02.2026

Valid until: 21.02.2027

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: CBM
Serial no: Yankee
Service interval: 12 months

Location: Yankee
Model: Oxybox
Build year: —

Cylinder size 5L	Hydro test 01.12.2018	Next hydro test 01.12.2028	Filled date —	Oxygen cleaned No
1st stage SN 07061	2nd stage SN —	Cylinder SN 81097020		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (99.9)	Pass
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Comment: changed	
		Flow test regulator	Pass

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

