

OXYBOX SERVICE CERTIFICATE

DSC-2026-000204

Service date: 22.01.2026

Valid until: 22.01.2027

Result: **Approved**

Equipment details

Customer: Kystarbeid AS
Producer: SANOSUB
Serial no: 064
Service interval: 12 months

Location: MS Nitrox
Model: Oxybox
Build year: 2024

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	01.05.2024	01.05.2034	—	-
1st stage SN	2nd stage SN	Cylinder SN		
—	—	—		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Comment: Freeflow mask changed	
Comment: Crimped new connector		Freeflow mask	-
Equipment cleaned	Pass	Flow test regulator	-

Notes

Oxybox in Green tube.

Technician

Serviced by: Daniel Kristensen

Signature

Electronically signed by DS Bergen AS

