

OXYBOX SERVICE CERTIFICATE

DSC-2026-000209

Service date: 22.01.2026

Valid until: 22.01.2027

Result: **Approved**

Equipment details

Customer: Valen marina AS
Producer: DAN medical
Serial no: 201906270019
Service interval: 12 months

Location: Valen Marina
Model: Oxybox
Build year: 2020

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
2L	01.10.2020	01.10.2030	—	-
1st stage SN	2nd stage SN	Cylinder SN		
—	—	—		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Comment: Freeflow mask changed	
Equipment cleaned	Pass	Freeflow mask	-
		Flow test regulator	-

Notes

Cylinder SN VDN003 UT
Reg SN E0003864

Technician

Serviced by: Daniel Kristensen

Signature

Electronically signed by DS Bergen AS

