

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000209

Service date: 22.01.2026

Valid until: 22.01.2027

Result: **Approved**

## Equipment details

Customer: Valen marina AS  
Producer: DAN medical  
Serial no: 201906270019  
Service interval: 12 months

Location: Valen Marina  
Model: Oxybox  
Build year: 2020

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
2L	01.10.2020	01.10.2030	—	-
1st stage SN	2nd stage SN	Cylinder SN		
—	—	—		

## Service items

General condition	Pass	Equipment leak test	Pass
O <sub>2</sub> regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Comment: Freeflow mask changed	
Equipment cleaned	Pass	Freeflow mask	-
		Flow test regulator	-

## Notes

Cylinder SN VDN003 UT  
Reg SN E0003864

## Technician

Serviced by: Daniel Kristensen

## Signature

Electronically signed by DS Bergen AS

