

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000214

Service date: 22.01.2026

Valid until: 22.01.2027

Result: **Approved**

## Equipment details

Customer: IMC Diving AS  
Producer: CBM  
Serial no: 737  
Service interval: 12 months

Location: 737  
Model: Oxybox  
Build year: 2012

| Cylinder size          | Hydro test        | Next hydro test       | Filled date | Oxygen cleaned |
|------------------------|-------------------|-----------------------|-------------|----------------|
| 5L                     | 01.08.2024        | 24.08.2034            | —           | -              |
| 1st stage SN<br>407302 | 2nd stage SN<br>— | Cylinder SN<br>J63083 |             |                |

## Service items

|                          |      |                                |      |
|--------------------------|------|--------------------------------|------|
| General condition        | Pass | Equipment leak test            | Pass |
| O <sub>2</sub> regulator | Pass | Oxygen analysis (-)            | -    |
| Gauge                    | Pass | Pocket mask                    | Pass |
| Hose                     | Pass | Comment: Freeflow mask changed |      |
| Comment: Changed         |      | Freeflow mask                  | Pass |
| Equipment cleaned        | Pass | Flow test regulator            | Pass |

## Notes

Oxybox yellow case  
regulator SN 407302  
Cylinder SN J63083

## Technician

Serviced by: Daniel Kristensen

Signature

