

Equipment details

Customer: Dypvest AS

Producer: SANOSUB

Serial no: 024 UT

Service interval: 12 months

Location: -

Model: Oxybox

Build year: 2025

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	01.03.2025	01.03.2035	01.07.2025	Yes (01.07.2025)
1st stage SN	2nd stage SN	Cylinder SN		
1370	—	024 UT		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (99.9)	Pass
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Comment: Changed	
		Flow test regulator	Pass

Technician

Serviced by: Daniel Kristensen

Signature



