

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000295

Service date: 04.02.2026

Valid until: 04.02.2027

Result: **Approved**

## Equipment details

Customer: Havforskningsinstituttet  
Producer: SANOSUB  
Serial no: NA  
Service interval: 12 months

Location: AVD Bergen  
Model: Oxybox  
Build year: 2022

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	01.11.2022	01.11.2032	—	-
1st stage SN 12221	2nd stage SN —	Cylinder SN MLI046UT		

## Service items

General condition	Pass	Equipment leak test	Pass
O <sub>2</sub> regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Comment: Changed Flow test regulator	Pass

## Notes

Filled up to 200Bar with oxygen

## Technician

Serviced by: Daniel Kristensen

Signature

*Daniel Kristensen*

