

# INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000346

Service date: 18.02.2026

Valid until: 18.02.2027

Result: **Approved**

## Equipment details

Customer: Otech AS  
Producer: Novasub  
Serial no: 371129013  
Customer tag: Diver 2

Location: Transporter  
Model: Pneumo 0-60M  
Build year: 2024  
Service interval: 12 months

## Calibration Instrument

Input basis: msw

Accuracy (%FS): 0.25

Range: 0 + 60 m

Bar → msw factor: 1

Full scale: 60 msw

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +/-2°C Traceability to international standards

## Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	5.01	0.010	0.0167	PASS
2	10.0	10.01	0.010	0.0167	PASS
3	20.0	20.026	0.026	0.0433	PASS
4	30.0	30.023	0.023	0.0383	PASS
5	40.0	40.033	0.033	0.0550	PASS
6	50.0	50.03	0.030	0.0500	PASS

## Technician

Serviced by: Daniel Kristensen

Signature

*Daniel Kristensen*

