

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000373

Service date: 21.02.2026

Valid until: 21.02.2027

Result: **Approved**

## Equipment details

Customer: Sjøentreprenøren AS  
Producer: CBM  
Serial no: Yankee  
Service interval: 12 months

Location: Yankee  
Model: Oxybox  
Build year: —

| Cylinder size         | Hydro test        | Next hydro test         | Filled date | Oxygen cleaned |
|-----------------------|-------------------|-------------------------|-------------|----------------|
| 5L                    | 01.12.2018        | 01.12.2028              | —           | No             |
| 1st stage SN<br>07061 | 2nd stage SN<br>— | Cylinder SN<br>81097020 |             |                |

## Service items

|                          |      |                        |      |
|--------------------------|------|------------------------|------|
| General condition        | Pass | Equipment leak test    | Pass |
| O <sub>2</sub> regulator | Pass | Oxygen analysis (99.9) | Pass |
| Gauge                    | Pass | Pocket mask            | Pass |
| Hose                     | Pass | Freeflow mask          | Pass |
| Equipment cleaned        | Pass | Comment: changed       |      |
|                          |      | Flow test regulator    | Pass |

## Technician

Serviced by: Daniel Kristensen

Signature

*Daniel Kristensen*