

OXYBOX SERVICE CERTIFICATE

DSC-2026-000380

Service date: 22.02.2026

Valid until: 22.02.2027

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: CBM
Serial no: 080172917A
Service interval: 12 months

Location: Mjø
Model: Oxybox
Build year: —

Cylinder size 5L	Hydro test 01.11.2017	Next hydro test 01.11.2027	Filled date 01.03.2024	Oxygen cleaned No
1st stage SN 210568081	2nd stage SN 20061214	Cylinder SN AIL137UT		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (99.9)	Pass
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Comment: changed	
		Flow test regulator	Pass

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

