

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000424

Service date: 23.02.2026

Valid until: 23.02.2027

Result: **Approved**

Equipment details

Customer: Dypvest AS

Producer: Novasub

Serial no: 188650007

Customer tag: Divepanel SN 2502 Diver 2

Location: Båt

Model: Pneumo 0-60M

Build year: 2024

Service interval: 12 months

Calibration Instrument

Input basis: msw

Accuracy (%FS): 0.25

Range: 0 + 70 m

Bar → msw factor: 9.94

Full scale: 70 msw

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

| # | Instrument (msw) | Reference (bar) | Dev (msw) | %FS | Result |
|---|------------------|-----------------|-----------|--------|--------|
| 1 | 5.0 | 0.501 | 0.0201 | 0.0287 | PASS |
| 2 | 10.0 | 1.007 | 0.0096 | 0.0137 | PASS |
| 3 | 20.0 | 2.005 | 0.0703 | 0.1004 | PASS |
| 4 | 30.0 | 3.005 | 0.1303 | 0.1861 | PASS |
| 5 | 40.0 | 4.009 | 0.1505 | 0.2150 | PASS |
| 6 | 50.0 | 5.04 | 0.0976 | 0.1394 | PASS |

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

