

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000456

Service date: 23.02.2026

Valid until: 23.02.2027

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: Gauge
Serial no: 21092010 Pneumo D2
Customer tag: –

Location: Mjø
Model: Pneumo
Build year: –
Service interval: 12 months

Calibration Instrument

Input basis: msw Range: 0 + 70 m Full scale: 70 msw
Accuracy (%FS): 0.25 Bar → msw factor: 9.94

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	0.501	0.0201	0.0287	PASS
2	10.0	1.012	0.0593	0.0847	PASS
3	20.0	2.01	0.0206	0.0294	PASS
4	30.0	3.01	0.0806	0.1151	PASS
5	40.0	4.018	0.0611	0.0873	PASS
6	50.0	5.015	0.1509	0.2156	PASS

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

