

# INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000167

Service date: 12.01.2026

Valid until: 12.01.2027

Result: **Approved**

## Equipment details

Customer: Dypvest AS  
Producer: AXSUB  
Serial no: 002-0093  
Customer tag: –

Location: Båt  
Model: Depthsensor  
Build year: –  
Service interval: 12 months

## Calibration Instrument

Input basis: msw  
Accuracy (%FS): 1

Range: 0 + 70 m  
Bar → msw factor: 1

Full scale: 70 msw

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS  
Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

## Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	4.9511	0.0489	0.0699	PASS
2	10.0	9.79	0.210	0.3000	PASS
3	20.0	19.85	0.150	0.2143	PASS
4	30.0	29.66	0.340	0.4857	PASS
5	40.0	39.54	0.460	0.6571	PASS
6	50.0	49.54	0.460	0.6571	PASS

## Technician

Serviced by: Daniel

Signature

*Daniel K. Sørensen*

