

OXYBOX SERVICE CERTIFICATE

DSC-2026-000524

Service date: 06.03.2026

Valid until: 06.03.2027

Result: **Approved**

Equipment details

Customer: IMC Diving AS
Producer: CBM
Serial no: 714
Service interval: 12 months

Location: 714
Model: Oxybox
Build year: —

| | | | | |
|----------------------------|--------------------------|-------------------------------|---------------------------|------------------------------------|
| Cylinder size 5L | Hydro test 19.02.2025 | Next hydro test 19.02.2035 | Filled date 01.03.2025 | Oxygen cleaned Yes (01.03.2025) |
| 1st stage SN 081023225A | 2nd stage SN 20080109 | Cylinder SN AA76058 | | |

Service items

| | | | |
|--------------------------|------|-------------------------|------|
| General condition | Pass | Equipment leak test | Pass |
| O ₂ regulator | Pass | Oxygen analysis (100.0) | Pass |
| Gauge | Pass | Pocket mask | Pass |
| Hose | Pass | Freeflow mask | Pass |
| Equipment cleaned | Pass | Flow test regulator | Pass |

Technician

Serviced by: Daniel Kristensen

Signature

