

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000566

Service date: 18.03.2026

Valid until: 18.03.2027

Result: **Approved**

Equipment details

Customer: Valen marina AS
Producer: 3D instruments
Serial no: 99839-37-19
Customer tag: Diver 2 Red

Location: Poseidon
Model: Pneumo Gauge
Build year: 2021
Service interval: 12 months

Calibration Instrument

Input basis: msw
Accuracy (%FS): 0.25

Range: 0 + 60 m
Bar → msw factor: 10

Full scale: 60 msw

Calibration Instrument: —
Calibration Instrument verified against: —

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	0.511	0.110	0.1833	PASS
2	10.0	1.032	0.320	0.5333	FAIL
3	20.0	2.021	0.210	0.3500	FAIL
4	30.0	3.012	0.120	0.2000	PASS
5	50.0	5.023	0.230	0.3833	FAIL

Remarks

Not in use to measure depth

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

