

OXYBOX SERVICE CERTIFICATE

DSC-2026-000207

Service date: 22.01.2026

Valid until: 22.01.2027

Result: **Approved**

Equipment details

Customer: Valen marina AS
Producer: SANOSUB
Serial no: 03
Service interval: 12 months

Location: DS3
Model: Oxybox
Build year: 2019

Cylinder size	Hydro test	Next hydro test	Filled date	Oxygen cleaned
5L	01.06.2019	01.06.2029	—	-
1st stage SN	2nd stage SN	Cylinder SN		
—	—	—		

Service items

General condition	Pass	Equipment leak test	Pass
O ₂ regulator	Pass	Oxygen analysis (-)	-
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Comment: Freeflow mask changed	
Comment: Changed		Freeflow mask	Pass
Equipment cleaned	Pass	Flow test regulator	Pass

Notes

Cylinder SN EED 146UT

Technician

Serviced by: Daniel Kristensen

Signature

Daniel Kristensen

