

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000600

Service date: 06.04.2026

Valid until: 06.04.2027

Result: **Approved**

## Equipment details

Customer: Sjøentreprenøren AS  
Producer: CBM  
Serial no: 1056  
Service interval: 12 months

Location: Alu kontainer  
Model: Oxybox  
Build year: —

Cylinder size 5L	Hydro test 01.03.2025	Next hydro test 01.03.2035	Filled date 06.04.2026	Oxygen cleaned Yes (06.04.2026)
1st stage SN 0803833378A	2nd stage SN 20070808	Cylinder SN 25AG062 021UT		

## Service items

General condition	Pass	Equipment leak test	Pass
O <sub>2</sub> regulator	Pass	Oxygen analysis (99.6)	Pass
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Flow test regulator	Pass

## Notes

New cylinder

## Technician

Serviced by: Daniel Kristensen

Signature

