

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000611

Service date: 06.04.2026

Valid until: 06.04.2027

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: Wika
Serial no: 1A00TWTFI3Q
Customer tag: Pneumo D1

Location: Alu kontainer
Model: 0-6 Bar Digital
Build year: 2018
Service interval: 12 months

Calibration Instrument

Input basis: bar

Range: 6 bar

Full scale: 6 bar

Accuracy (%FS): 0.1

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +/-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (bar)	Reference (bar)	Dev (bar)	%FS	Result
1	1.0	0.9958	0.0042	0.0700	PASS
2	2.0	1.9965	0.0035	0.0583	PASS
3	3.0	2.996	0.004	0.0667	PASS
4	4.0	3.9952	0.0048	0.0800	PASS
5	5.0	4.9956	0.0044	0.0733	PASS

Technician

Serviced by: Daniel Kristensen

Signature

