

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000612

Service date: 06.04.2026

Valid until: 06.04.2027

Result: **Approved**

Equipment details

Customer: Sjøentreprenøren AS
Producer: Wika
Serial no: 1A00TWTGH3Q
Customer tag: Pneumo D2

Location: Alu kontainer
Model: 0-6 Bar Digital
Build year: 2018
Service interval: 12 months

Calibration Instrument

Input basis: bar

Range: 6 bar

Full scale: 6 bar

Accuracy (%FS): 0.1

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS
Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +/-2°C Traceability to international standards

Verification results (As-Found)

| # | Instrument (bar) | Reference (bar) | Dev (bar) | %FS | Result |
|---|------------------|-----------------|-----------|--------|--------|
| 1 | 1.0 | 0.9998 | 0.0002 | 0.0033 | PASS |
| 2 | 2.0 | 1.9994 | 0.0006 | 0.0100 | PASS |
| 3 | 3.0 | 2.9996 | 0.0004 | 0.0067 | PASS |
| 4 | 4.0 | 3.9994 | 0.0006 | 0.0100 | PASS |
| 5 | 5.0 | 4.9989 | 0.0011 | 0.0183 | PASS |

Technician

Serviced by: Daniel Kristensen

Signature

