

# HYPERBARIC CHAMBER SERVICE CERTIFICATE

DSC-2026-000615

Service date: 08.04.2026

Valid to: 08.04.2027

Result: **Approved**

## Equipment details

Customer: Valen marina AS  
Producer: Draeger  
Serial no: 1244  
Service interval: 12 months

Location: Chamber Container  
Model: Chamber  
Build year: 1982

## Pressure

Pressure reducing valve	—	Pass
Main chamber overpressure valve	5.5	Pass
Entry lock overpressure valve	—	-
Pipe maximum working pressure	300.0	Pass

Pressure test  
Pressure: — bar  
Time: — min  
Descent: — min  
Ascent: 0 min

Hydro  
Last hydrostatic test:  
30.08.2023  
Next hydrostatic test:  
30.08.2033  
Hydrotested: No  
Testrange: —

## Checklist

General condition	Pass	Communication – main chamber	Pass
Entry lock inlet valve	Pass	Communication – Entry lock	Pass
Entry lock outlet valve	Pass	Gasket – medical lock	Pass
Main chamber inlet valve	Pass	Gasket – Entry lock door	Pass
Main chamber outlet valve	Pass	Gasket – main chamber door	Pass
Depth alarm – audible signal DE-OX	Pass	Window/viewport 06/2023	Pass
Depth gauge	Pass	Valves & Piping	Pass
High pressure gauge (HP)	Pass	Lightening	Pass
Low pressure gauge (LP)	Pass	Cameras	Pass
BIBS / O <sub>2</sub> mask see bibs certificate	Pass	Surface treatment inside/outside	Pass
LP bank	-	Equalizer	Pass
HP Bank Composite 8x50L 200Bar	Pass	Drainage	Pass
O <sub>2</sub> needle valves internal / external	Pass		

## Technician

Serviced by: Daniel Kristensen

Signature

