

INSTRUMENT VERIFICATION CERTIFICATE

DSC-2026-000620

Service date: 08.04.2026

Valid until: 08.04.2027

Result: **Approved**

Equipment details

Customer: Valen marina AS
Producer: Gauge
Serial no: Main Chamber
Customer tag: Mainlock panel

Location: Chamber Container
Model: Pneumo
Build year: —
Service interval: 12 months

Calibration Instrument

Input basis: msw

Range: 0 + 60 m

Full scale: 60 msw

Accuracy (%FS): 0.25

Bar → msw factor: 10

Calibration Instrument: AMETEK XP2I Digital pressure calibrator Certified range 0 to 20 / 0-5000PSI Bar, SN no: 404853 Certificate no: 451333 Date of issue: 24.06.2025 Uncertainty of computed pressure: +/- 0,02 % FS Temperature: 20°C

Calibration Instrument verified against: Fluke PPC4 829 300psi Pressure controller Accuracy: 0,01 % reading Serial no: 829 Certificate no: 24001. Calibration temp 20°C +-2°C Traceability to international standards

Verification results (As-Found)

#	Instrument (msw)	Reference (bar)	Dev (msw)	%FS	Result
1	5.0	0.502	0.020	0.0333	PASS
2	10.0	1.008	0.080	0.1333	PASS
3	20.0	2.0087	0.087	0.1450	PASS
4	30.0	3.0055	0.055	0.0917	PASS
5	50.0	5.0038	0.038	0.0633	PASS

Technician

Serviced by: Daniel Kristensen

Signature

