

HYPERBARIC CHAMBER SERVICE CERTIFICATE

DSC-2026-000651

Service date: 13.04.2026

Valid to: 13.04.2027

Result: **Approved**

Equipment details

Customer: Kystarbeid AS
Producer: Imenco
Serial no: 1385
Service interval: 12 months

Location: Chamber Container
Model: Chamber 1800MM
Build year: 2010

Pressure

Pressure reducing valve	15.0	Pass	Pressure test	Hydro
Main chamber overpressure valve	6.3	Pass	Pressure: – bar	Last hydrostatic test:
Recommend install valve to emergency close			Time: – min	01.06.2022
Entry lock overpressure valve	–	-	Descent: – min	Next hydrostatic test:
Pipe maximum working pressure	200.0	Pass	Ascent: – min	01.06.2032
				Hydrotested: No
				Testrange: –

Checklist

General condition	Pass	Communication – main chamber	Pass
Entry lock inlet valve	Pass	Communication – Entry lock	Pass
Entry lock outlet valve	Pass	Gasket – medical lock	Pass
Main chamber inlet valve	Pass	Gasket – Entry lock door	Pass
Main chamber outlet valve	Pass	Gasket – main chamber door	Pass
Depth alarm – audible signal	Pass	Window/viewport	Pass
Depth gauge	Pass	Valves & Piping	Pass
see calibration certificate		Lightening	Pass
High pressure gauge (HP)	Pass	Cameras	Pass
Low pressure gauge (LP)	Pass	Surface treatment inside/outside	Pass
BIBS / O ₂ mask	Pass	Equalizer	Pass
see bibs certificate, Recommend to install valves inlet and outlet on block		Drainage	Pass
LP bank	-		
HP Bank	Pass		
3x80L 300bar, New 2024			
O ₂ needle valves internal / external	Pass		

Technician

Serviced by: Daniel Kristensen

Signature

