

# OXYGEN REGULATOR

Certificate no: DSC-2026-000656

Service date: 15.04.2026

Valid until: 15.04.2027

Result: **Approved**

## Equipment details

Customer: Kystarbeid AS  
Producer: Zenon Regulator  
Serial no: KA2014  
Service interval: 12 months

Location: Chamber Container  
Model: Oxygen regulator  
Build year: -

## NOTE

Oxygen regulator ZENON 0-10Bar Tested and approved.  
OPV opens at 15.5 Bar

Oxygen cleaned and ready to use

## Technician

Serviced by: Daniel Kristensen

Signature

