

DIVATOR MKII SET

Certificate no: DSC-2026-000680

Service date: 30.04.2026

Valid until: 30.04.2027

Result: **Approved**

Equipment details

Customer: IMC Diving AS

Producer: Interspiro

Serial no: M2252109 & B2145106

Service interval: 12 months

Location: 105

Model: Divator MKII

Build year: -

NOTE

Tested with Labtec PRO

This is a certificate confirmation, the documentation has a separate file.

Technician


Serviced by: Daniel Kristensen

Signature



test report

test report: 468
 device no.: M2252109
 type of Yearly test
 test system: PRO 3102.24.325

| | | | |
|---|---------------------|---------------------|------------------------------|
| workshop: DS BERGEN AS | | person: | Daniel Kristensen |
| street: Storebotn 55G | | phone: | 91649162 |
| place: 5309 Kleppestø | | fax: | |
| | | email: | Daniel@dsbergen.no |
| customer: IMC Diving AS | | person: | Jarle |
| street: Kristiansholm 8 | | phone: | |
| place: 5035 Bergen | | fax: | |
| | | email: | jarle.johansen@imc-diving.no |
| device data: | deadlines: | device-info: | service: |
| device no: M2252109 | 1 y test 30.04.2027 | | - 1 year service |
|  | 1 y srv 30.04.2027 | | - cleaning and disinfection |
| | | | |
| type: Interspiro Divator PP | | | |
| class: mask PP | | | |
| location: 105 | | | |

divator sett with Kirby block.

| tests/measurements | unit | min | max | actual | result |
|--|------|-----|-----|--------|--------|
| visual test (0 = ok) input mask: (0 = classic, > 0 = extended/list...) | | | 0 | 0 | OK |
| mask fitted approx. 6 sec head expansion Mask volume 0,2 l | | | | 0 | OK |
| leakage test w/ open exhalation valve 0,25 min at p=-10 mbar | mbar | | 7,5 | 2,1 | OK |

test executed: 30.04.2026 10:47
 by: Daniel Kristensen

test signed: 30.04.2026 10:47
 by: Daniel Kristensen

LabNet V: 2025.12.2.0

next test on: 30.04.2027

result:

device ready for use

test report

test report: 468
 device no.: B2145106
 type of Yearly test
 test system: PRO 3102.24.325

| | | | |
|-------------------------------------|---|---------------------|---|
| workshop: DS BERGEN AS | | person: | Daniel Kristensen |
| street: Storebotn 55G | | phone: | 91649162 |
| place: 5309 Kleppestø | | fax: | |
| | | email: | Daniel@dsbergen.no |
| customer: IMC Diving AS | | person: | Jarle |
| street: Kristiansholm 8 | | phone: | |
| place: 5035 Bergen | | fax: | |
| | | email: | jarle.johansen@imc-diving.no |
| device data: | deadlines: | device-info: | service: |
| device no: B2145106 | 1 y test 30.04.2027 1 y srv 30.04.2027 | | 1 year service cleaning and disinfection |
| type: Interspiro BV Divator PP | | | |
| class: demand valve overpressure | | | |
| location: 105 | | | |

divator sett with Kirby block.

| tests/measurements | unit | min | max | actual | result |
|---|------|-----|------|--------|--------|
| visual test (0 = ok) input mask: (0 = classic, > 0 = extended/list...) | | | 0 | 0 | OK |
| demand valve activation pressure at 10 l/min suction | mbar | -25 | -0,1 | -6,8 | OK |
| static positive pressure at bar cylinder pressure, suction 0 0=yes 1=no | mbar | 2 | 5 | 4,6 | OK |
| inhalation resistance at 40 x 2,5 l/min 5 hubs HP .>. bar | mbar | 0,2 | 10 | 2,2 | OK |
| pressure release of test equipment (HP, MP and LP) | | | | 0 | OK |

test executed: 30.04.2026 10:47
 by: Daniel Kristensen
 test signed: 30.04.2026 10:47
 by: Daniel Kristensen

result:
device ready for use

LabNet V: 2025.12.2.0

next test on: 30.04.2027