

# OXYBOX SERVICE CERTIFICATE

DSC-2026-000712

Service date: 27.05.2026

Valid until: 27.05.2027

Result: **Approved**

## Equipment details

Customer: Dypvest AS

Producer: SANOSUB

Serial no: 1083

Service interval: 12 months

Location: Båt

Model: Oxybox

Build year: 2021

Cylinder size 5L	Hydro test 01.06.2021	Next hydro test 01.06.2031	Filled date 27.05.2026	Oxygen cleaned No
1st stage SN 041	2nd stage SN 487	Cylinder SN L45216		

## Service items

General condition	Pass	Equipment leak test	Pass
O <sub>2</sub> regulator	Pass	Oxygen analysis (99.9)	Pass
Gauge	Pass	Pocket mask	Pass
Hose	Pass	Freeflow mask	Pass
Equipment cleaned	Pass	Comment: Changed	
		Flow test regulator	Pass

## Technician

Serviced by: Daniel Kristensen

Signature

