

# HYPERBARIC CHAMBER SERVICE CERTIFICATE

DSC-2026-000261

Service date: 28.01.2026

Valid to: 20.10.2026

Result: **Approved**

## Equipment details

Customer: Amundsen Diving AS  
Producer: Haux  
Serial no: 200604  
Service interval: 12 months

Location: Scuba  
Model: 1800MM  
Build year: 2006

## Pressure

Pressure reducing valve	20.0	Pass	Pressure test	Hydro
Main chamber overpressure valve	4.7	Pass	Pressure: – bar	Last hydrostatic test: 20.10.2016
Entry lock overpressure valve	–	Pass	Time: – min	Next hydrostatic test: 20.10.2026
Pipe maximum working pressure	–	Pass	Descent: – min	Hydrotested: No
			Ascent: – min	Testrange: –

## Checklist

General condition	Pass	Communication – main chamber	Pass
Entry lock inlet valve	Pass	Communication – Entry lock	Pass
Entry lock outlet valve	Pass	Gasket – medical lock	Pass
Main chamber inlet valve	Pass	Gasket – Entry lock door	Pass
Main chamber outlet valve	Pass	Gasket – main chamber door	Pass
Depth alarm – audible signal	-	Window/viewport	Pass
Depth gauge	Pass	New 2017	
Calibrated		Valves & Piping	Pass
High pressure gauge (HP)	Pass	Lightening	Pass
Low pressure gauge (LP)	Pass	Cameras	-
BIBS / O <sub>2</sub> mask	Pass	Surface treatment inside/outside	Pass
LP bank	-	Equalizer	Pass
HP Bank	Pass	Drainage	Pass
O <sub>2</sub> needle valves internal / external	Pass		

## Technician

Serviced by: Daniel Kristensen

Signature

*Daniel Kristensen*

